

The logo for DNW, featuring a stylized blue 'D' with a small black square detail on its left side, followed by the letters 'N' and 'W' in a bold, blue, sans-serif font.

DNW

PROTECTIVE MASK

GB2626-2006 KN95



EU 149:2001 +A1:2009 · FFP2 NR D

FFP2

DNW

PROTECTIVE MASK

GB2626-2006 (Civil Grade)

KN95

APPROVED



Antiviral fiber



Skin-friendly



Particulate matter
filtration



individual package



FDA

CE

EU 149:2001 +A1:2009 • FFP2 NR D

Standard size/Earloop



**Individual Package
Clean & Hygienic**

Product Name: DUONUOWAN Protective Masks(Civil Grade)

Main Material: Nonwoven Fabric, Meltblown Fabric,
Needle-punched Cotton

Execution Standard: GB2626-2006 / EU 149:2001 +A1:2009

Expiration Date: 3 Years

See the packaging indication for production batch number
and expiration date.

PROTECTION LEVEL

KN95

FFP2


N95



EU 149:2001 +A1:2009 · FFP2 NR D

Product Parameters

DNW



SPECIFICATIONS

MASK

(Independent Packages)

Size:10.7x16cm




40 pcs/Box
Size:16x13x23cm




10 pcs/Box
Size:12.8x9x16cm


New & Good Materials, No Secondary Pollution



Multi-effect Protection & Filtration



How to Use



1. Open the mask to make the nose clip at the top, and pull the ear straps with both your hands.



2. Hold the mask against your chin to completely cover your nose and mouth.



3. Pull the ear straps behind your ears and adjust them to make you feel comfortable.



4. Use both your hands to adjust the shape of the nose clip. Place your fingers in the middle of the nose clip and press it inwards while moving your fingertips along both sides of the nose clip until it is pressed to fit the bridge of your nose.
(Making the nose clip get a seal with only one hand may affect the tightness of the mask).



5. Cover the mask with your hand and exhale vigorously. If you feel the air escaping from the nose clip, it is required to tighten the nose clip; if the air escapes from the edge of the mask, readjust the headband to ensure tightness.

Inspection Method

Before use, check whether the package of the mask is damaged and the components are intact. In case of any stain or damage, discontinue use immediately. Before entering the work area, it is required to check the air tightness of the mask. If you feel a leak, please wear it again according to the instructions for how to wear a mask. You can enter the work area only after feeling no leak. (If you feel the respiratory resistance is obviously increased or the mask falls short of the standard for use, it is recommended to replace the mask immediately.)

KN95

GB2626-2006
(Civil Grade)

FFP2

N95

This dust mask is used to prevent non-oily suspended particles.
Filtration Efficiency $\geq 95\%$.



Do not use if package is damaged!
Please refer to the manual before use and use a mask correctly!







Fiscal Year 2020

CERTIFICATION OF FDA REGISTRATION

This certifies that:

has completed the FDA Establishment Registration and Device Listing with the US Food & Drug Administration, through

Shenzhen CCT Testing Technology Co., Ltd.



Owner/Operator Number: 10062783


Device Listing#:

Listing No	Code	Device Name	Proprietary Name
D373895	KHA	MASK, SCAVENGING	ACCESSORY, SURGICAL APPAREL (FACE MASK , DISPOSABLE FACE MASK, SURGICAL MASK, DISPOSABLE SURGICAL FACE MASK)

CCT will confirm that such registration remains effective upon request and presentation of this certificate until the end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. CCT makes no other representations or warranties, nor does this certificate make any representations or warranties to any person or entity other than the named certificate holder, for whose sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. CCT assumes no liability to any person or entity in connection with the foregoing.

Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding." The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration, CCT is not affiliated with the U.S. Food and Drug Administration.

Shenzhen CCT Testing Technology Co., Ltd.
W: www.fda-test.com E: fda@fda-test.com
T: **400-8788-298** T: 86-755-36916737


Chief engineer
Issued: 03/07/2020
Expiration Date: 12/31/2020



Web: <http://www.fda.gov> Tel: 1-888-INFO-FDA (1-888-463-6332) e-mail: webmail@oc.fda.gov

Certificate of Compliance



No. 3J200303.GQT0S72

Certificate's
Holder:

Certification ECM
Mark:



Product:
Model(s):

Protective masks
Protective mask KN95,
Disposable mask 175mmx95mm
(Class of device: FFP2 NR D)

Verification to:

Standard:
EN 149:2001+A1:2009

related to CE Directive(s):
R 2016/425 (Personal Protective Equipment)

Remark: The product(s) has been verified on a voluntary basis. The product(s) satisfies the requirements of the Certification Mark of ECM, in reference to the above listed Standard(s). The above Compliance Mark can be affixed on the product(s) accordingly to the ECM regulation about its release and its use. The regulation can be found at www.entecerma.it. This Certificate of Compliance can be checked for validity at www.entecerma.it.

This verification doesn't imply assessment of the production of the product(s).

Additional information, clarification about the **CE** Marking:



We attest that a TCF for the **CE** Marking process is in place. Whereas the Manufacturer is Responsible to start the **CE Marking Certification Procedure** through an appointed Notified Body and the perform all the necessary activities, as required by the Directive and accepted by the Notified Body, before placing the **CE** Mark on the product(s).

Date of issue 03 March 2020

Expiry date 02 March 2025

Chief Manager
Marco Morino

Deputy Manager
Amanda Payne

Ente Certificazione Macchine Srl

Via Ca' Bella, 243 - Loc. Castello di Serravalle - 40053 Valsamoggia (BO) - ITALY
☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it



中国认可
国际互认
检测
TESTING
CNAS L0823



201719001121

广州市微生物研究所

GUANG ZHOU INSTITUTE OF MICROBIOLOGY

检测报告

TEST REPORT

报告编号

KJ20200309

样品名称

多诺万防护口罩

委托单位

广东乾景检测有限公司





中国认可
国际互认
检测
TESTING
CNAS L0823



检测编号: KJ20200309
Test No.

广州市微生物研究所
GUANG ZHOU INSTITUTE OF MICROBIOLOGY

检测报告
TEST REPORT

收样日期: 2020 年 02 月 15 日
Date Received

检测日期: 2020 年 02 月 25 日
Date Analyzed

样品名称 Name of Sample	多诺万防护口罩	检测类型 Test Type	委托检测
委托单位 Applicant	广东乾景检测有限公司	样品来源 Sample Source	委托方送检
生产单位 Manufacturer	广东乾景检测有限公司	商标 Brand	
生产日期 Date of Production	2020.02.14	生产批号 Batch Number	20200213A
型号规格 Type and Specification	KN95	样品数量 Quantity of Sample	1 份 (55 片)
样品状态 State of Sample	片状	样品包装 Packing of Sample	袋装
样品图片 Sample Picture			
检验依据和方法 Standard and Methods	GB 2626-2006 呼吸防护用品 自吸过滤式防颗粒物呼吸器		
检测项目 Items of Analysis	1. 过滤效率 2. 呼气阻力 3. 吸气阻力 4. 死腔 5. 头带 6. 阻燃性		
备注 Remarks	——		

接下页/To be continued



中国认可
国际互认
检测
TESTING
CNAS L0823



检测编号: KJ20200309
Test No.

广州市微生物研究所
GUANG ZHOU INSTITUTE OF MICROBIOLOGY

检测报告
TEST REPORT

收样日期: 2020 年 02 月 15 日
Date Received

检测日期: 2020 年 02 月 25 日
Date Analyzed

检测结果
Test Results

序号	检测项目	单位符号	检测结果	标准要求
1	过滤效率	%	95.6	KN90 (≥90.0%) KN95 (≥95.0%) KN100 (≥99.97%)
2	呼气阻力	Pa	31.5	≤250
3	吸气阻力	Pa	41.7	≤350
4	死腔	[%]	平均值0.7	≤1.0
5	头带	/	合格	10N 10s 无断裂
6	可燃性	/	合格	预处理后燃烧不超过5s

报告结束/End of report

编制:
Editor

徐国洋

审核:
Checker

马尼

签发:
Issuer

李

签发日期 (公历)
Date Reported





中国认可
国际互认
检测
TESTING
CNAS L0823



201719001121

声 明

- 一、 本检测报告涂改增删无效，未加盖检测单位“检验检测专用章”无效，无相关责任人签名无效，复印件无效。
- 二、 对送检样品，报告中的样品信息由委托方声称，本单位不对其真实性负责；本检测报告仅对送检样品负责。
- 三、 对报告的异议应于报告签发之日起 15 个工作日内向本单位提出，逾期视为承认本报告。微生物检测不复检。
- 四、 本检测报告及我单位名称不得用于产品标签、广告、评优及商品宣传等。
- 五、 报告中标“*”项目为还未通过广东省资质认定和中国合格评定国家认可委员会认可的项目；标“#”为只通过中国合格评定国家认可委员会认可的项目；标“+”为只通过广东省资质认定的项目。
- 六、 报告中未取得广东省资质认定的项目，检测数据和结果仅作为科研、教学或内部质量控制之用。
- 七、 因报告中所用语言产生的歧义，以中文为准。



联系地址：广州市黄埔区科学城尖塔山路 1 号

检验地址：（与联系地址不同时填写此项）

邮政编码：510663

联系电话：（8620）61302671

网址：<http://www.ggtest.com.cn>